PEROXTESMO MI

en

Test paper for the detection of lactoperoxidase in milk

The enzyme lactoperoxidase is always present in raw milk. It will only be deactived by high temperature heating (UHT-milk) but not through normal pasteurization. So this test paper reacts with UHT-milk lactoperoxidase negative and with pasteurized milk lactoperoxidase positive.

Presentation:
Box of 100 test papers 15 x 15 mm.

Color reaction:
Peroxtesmo MI reacts with lactoperoxidase by a coloration of blue flecks on a white background.

Limit of sensitivity:
3 % raw milk in UHT-milk

Instructions for use:
Remove one test paper from the box. Immediately close the box. Apply one drop of milk on the test paper and within 2 minutes the prescence of lactoperoxidase will be detected by a blue fleck.

Comment:
Any color change after 2 minutes should be ignored.
The presence of Azidiol leads to a weaker color reaction.

Storage conditions:
Protect from direct sunlight and moisture. Store container in a cool and dry place (storage temperature below 20 °C).

MACHEREY-NAGEL GmbH & Co. KG · Neumann-Neander-Str. 6–8 · 52355 Düren · Germany
Tel.: +49 24 21 969-0 · Fax: +49 24 21 969-199 · info@mn-net.com · ***www.mn-net.com***



REF	Artikelnummer / Item number / Référence produit / Referencia / Artikelnummer / Numero di catalogo / Numer artykułu	LOT	Chargencode / Batch identification / Numéro de lot / Código de lote / Productienummer / Codice del lotto / Numer partii
	Achtung! / Warning! / Attention ! / ¡Atención! / Let op! / Attenzione! / Uwaga!		Verwendbar bis / Use by / À utiliser avant / Fecha de caducidad / Te gebruiken tot / Utilizzare entro (anno / mese) / Przydatność do użycia
	Temperaturbegrenzung / Permitted storage temperature range / Limites de température / Límites de temperatura / Temperatuurbegrenzing / Limite de temperatura / Ograniczenie temperatury		Trocken aufbewahren / Store in a dry place / Conserver au sec / Mantener seco / Op een droge plaats bewaren / Mantenere asciutto / Przechowywać w suchym miejscu
	Gebrauchsanweisung beachten / Please read instructions for use / Respecter les instructions d'utilisation / Obsérvense las instrucciones de uso / Lees de bijsluiter / Consultare le istruzioni per l'uso / Przestrzegać instrukcję użycia		Packung geschlossen halten / Keep container closed / Refermer la boîte / Mantenga el envase cerrado / Verpakking gesloten houden / Conservare la confezione chiusa / Przechowywać pojemnik szczelnie zamknięty
Σ	Ausreichend für <n> Prüfungen / Contains sufficient for <n> tests / Contenu suffisant pour <n> tests / Contenido suficiente para <n> tests / Voldoende voor <n> tests / Contenuto sufficiente per <n> test / Wystarczająco dla kontroli <n>		

Rev. 2016-06 PD 02811 / A028031 / 90627 / 0662